

## PRODUCT DESCRIPTION

**Programmable regulators** with Ethernet connection are designed to measure temperature, relative humidity and barometric pressure of air in non-aggressive environment, to alarm indication and to control of external devices. Three galvanic no isolated binary inputs are intended for detection of binary signals. Regulators are available in wall-mount version or with probe on a cable. For measuring temperature and relative humidity of compressed air is used type HxxxxP.

**Two output relays** functions can be set from regulator keyboard or from computer. Each relay can be assigned to one of measured, detected or computed value (dewpoint temperature, absolute humidity, specific humidity mixing ratio and specific enthalpy). Setting of delay, hysteresis, audible alarm is enabled for each relay. It is possible to control output relay status via Ethernet too. The instrument may send a warning message if the measured value goes out of adjusted limits. Devices are equipped with four button keyboard and LCD display.

**The formats of Ethernet communication** that are supported: www pages with user-design possibility, Modbus TCP protocol, SNMPv1 protocol and SOAP. For setting of all parameters you can use *TSensor* software (see [www.cometsystem.cz/software.htm](http://www.cometsystem.cz/software.htm)).

type *	measured values	version	mounting
<b>H0530</b>	T	ambient air	wall
<b>H4531</b>	T	external probe Pt1000/3850 ppm	wall
<b>H3530</b>	T + RH + CV	ambient air	wall
<b>H3531</b>	T + RH + CV	probe on a cable	wall
<b>H3531P</b>	T + RH + CV	probe on a cable – pressure up to 25 bars	wall
<b>H7530</b>	T + RH + P + CV	ambient air	wall
<b>H7531</b>	T + RH + P + CV	probe on a cable	wall

\* models marked HxxxxZ are custom - specified devices

T...temperature, RH...relative humidity, P...barometric pressure, CV...computed values

## INSTALLATION , OPERATION AND CONFIGURATION

The mounting holes and connection terminals are accessible after unscrewing the four screws in the corners of case and removing the lid. Devices have to be mounted on a flat surface to prevent its deformation. For binary sensors and external probe connection choose shielded cables (external diameter 4 to 6.5mm) with wire cross-section 0.14 to 1.5mm<sup>2</sup>. Maximum binary sensor and external probe cable length is 10 m. Insert attached plug into unused cable glands too. The all cables should be located as far as possible from potential interference sources. Pay attention to device mounting, because incorrect choice of working position or place of measuring could adversely affect accuracy and long-term stability of measured values.

Actual parameters settings of each relay can be displayed by pressing of „▲“ key. To change any parameter, press the „Set“ key, enter password (default 0000) and set required value. Then click on „Set“ and pressing „Esc“ key exit setup mode. To change the password and to set all other parameters (acoustic alarm, computed value selection etc.) is used *Extended setting mode* (see manual for devices at [www.cometsystem.cz/manuals.htm](http://www.cometsystem.cz/manuals.htm)).

For network device connection it is necessary to know new suitable IP address (you can get it automatically from DHCP server or from your network administrator) and to have *TSensor* software installed. After you connect external probe, Ethernet cable and power adapter, you run *TSensor* program, set the new IP address, configure the device in accordance with your requirements and finally store the settings. The default IP address of each device is preset to **192.168.1.213**.

Devices don't require special maintenance. We recommend you periodic calibration for measurement accuracy validation.

## ERROR STATES

Device continuously checks its state during operation and if an error appears, it is displayed relevant code: **Err 1** – measured or calculated value is over the upper limit, **Err 2** – measured or calculated value is below the lower limit or pressure measurement error occurred, **Err 0**, **Err 3** a **Err 4** – it is a serious error, please contact distributor of the device, **Err5**, **Err6** - there is problem with assigned value to output relay, **Err9** – inserted password is not valid.

## SAFETY INSTRUCTIONS



- Humidity and temperature sensors of the regulator can not be operate and store without a filter cap.
- Temperature and humidity sensors have not to be exposed to direct contact with water and other liquids.
- It is not recommended to use the humidity regulators for long time under condensation conditions.
- Take care when unscrewing the filter cap as the sensor element could be damaged.
- If the sensing probe of H3531P regulator is installed, make sure that measured area is without pressure.
- Use only the power adapter according to technical specifications and approved according to relevant standards.
- Don't connect or disconnect devices while power supply voltage is on.
- Installation, electrical connection and commissioning should be performed by qualified personnel only.
- Devices contain electronic components, it needs to liquidate them according to currently valid conditions.
- For more information, please use detailed manuals and other documentation which are available at [www.cometsystem.cz/manuals.htm](http://www.cometsystem.cz/manuals.htm) or [www.cometsystem.cz/software.htm](http://www.cometsystem.cz/software.htm)

# Technical specifications

Device type	H4531	H0530	H3530, H7530	H3531	H7531	H3531P
Common parameters	Supply voltage: 9 to 30Vdc Binary inputs: low level input voltage ... 0 to 0.5V, high level input voltage ... 3 to 30V, auxiliary power supply +U ... 9 to 30 Vdc / overall current max. 500 mA	Power consumption: ~1W + binary sensors Relay outputs: max. switching voltage ... 50V, max. switching current ... 2A, max. switching power ... 60VA	Relay outputs: max. switching voltage ... 50V, max. switching current ... 2A, max. switching power ... 60VA	Relay outputs: max. switching voltage ... 50V, max. switching current ... 2A, max. switching power ... 60VA	Relay outputs: max. switching voltage ... 50V, max. switching current ... 2A, max. switching power ... 60VA	Relay outputs: max. switching voltage ... 50V, max. switching current ... 2A, max. switching power ... 60VA
Temperature measuring range	-200 to 600°C	-30 to +80°C	-30 to +80°C	-30 to 105°C	-30 to 105°C	-30 to 105°C
Accuracy of temperature measurement	±0.2°C (without probe)	±0.4°C	±0.4°C	±0.4°C	±0.4°C	±0.4°C
Relative humidity (RH) measuring range	—	0 to 100 %RH	0 to 100 %RH	0 to 100 %RH	0 to 100 %RH	0 to 100 %RH
Accuracy of humidity measurement from 5 to 95 %RH at 23°C	—	±2.5 %RH	±2.5 %RH	±2.5 %RH	±2.5 %RH	±2.5 %RH
Barometric pressure measuring range	—	600 to 1100 hPa (H7530)	600 to 1100 hPa (H7530)	600 to 1100 hPa	600 to 1100 hPa	600 to 1100 hPa
Accuracy of barometric pressure measurement at 23°C	—	±1.3hPa (H7530)	±1.3hPa (H7530)	±1.3hPa	±1.3hPa	±1.3hPa
Other calculated humidity variables	—	yes	yes	yes	yes	yes
Recommended calibration interval	2 years	2 years	1 year	1 year	1 year	1 year
Protection class of the case with electronics	IP40	IP40	IP40	IP40	IP40	IP40
Protection class of the sensors cover	—	IP40	IP40	IP40	IP40	IP40
Temperature operating range of the case with electronics *	-30 to +80°C	-30 to +80°C	-30 to +80°C	-30 to +80°C	-30 to +80°C	-30 to +80°C
Temperature operating range of the sensing element (sensors)	—	-30 to +80°C	-30 to +80°C	-30 to +105°C	-30 to +105°C	-30 to +105°C
Humidity operating range	0 to 100%RH	0 to 100%RH	0 to 100%RH	0 to 100%RH	0 to 100%RH	0 to 100%RH
Mounting position	any position	sensor cover downwards	sensor cover downwards	any position **	any position **	any position **
Storage temperature range (0 to 100%RH, no condensation)	-30 to +80°C	-30 to +80°C	-30 to +80°C	-30 to +80°C	-30 až +80°C	-30 to +80°C
Electromagnetic compatibility according to	EN 61326-1	EN 61326-1	EN 61326-1	EN 61326-1	ČSN EN 61326-1	EN 61326-1
Weight	340 g	340 g	360 g	410 (450, 530) g	410 (450, 530) g	460 (500, 580) g
Dimensions [mm]						
Electrical wiring	<p>The diagram illustrates the electrical wiring for the device. It shows an Ethernet connection (1) to the device. Power is supplied to the device (2) and a relay (R45). The relay is connected to two other relays, Relay 1 and Relay 2. Relay 1 is connected to an external probe sensor, and Relay 2 is connected to another external probe sensor. The diagram also shows the connection of binary inputs (+U, IN, GND, shield) to the device power supply. The external probe sensor is connected to the device through a shielded cable (P11000).</p>					

\* It is recommended to switch off the LCD display at ambient temperature above 70°C.

\*\* If it can lead to long term condensation of water, it is necessary to use the probe at position with sensor cover downwards